

Certificate



Certificate no.

CU 72228652 01

License Holder:

Heidolph Instruments
GmbH & Co. KG
Walpersdorfer Str. 12
91126 Schwabach
Deutschland

Manufacturing Plant:

--DE22ZTJM 001

Test report no.: USA- DE22ZTJM 002

Client Reference: Walter Lohmann

Tested to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Certified Product: Vacuum Pumps for Laboratory Use

License Fee - Units

Model :	Hei-VAC Valve Control;	(1)	7
Designation :	Hei-VAC Valve Industrial	(2)	
Input ratings :	(1) 230V 50/60Hz 1.8A;	or	
	(1) 100-115V 50/60Hz 3.4A / 120V 60Hz 3.4A;		
	(2) 100-115V 50/60Hz 5.7A /		
	200-230V 50/60Hz 3.0A		
Protection:	Class I; IP 40/Type 1(UL50E)		

7

Licensed Test mark:

Date of Issue

(day/mo/yr)

05/12/2022

