




America

CERTIFICATE

No. U8 069850 0015 Rev. 00

Holder of Certificate: **Heidolph Instruments GmbH & Co. KG**
 Walpersdorfer Str. 12
 91126 Schwabach
 GERMANY

Certification Mark:



Product: **Laboratory Equipment**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713260215

Date, 2022-08-31

(Benedikt Pulver)



America

CERTIFICATE

No. U8 069850 0015 Rev. 00

Model(s): Hei-VAP Value, Hei-VAP Value Digital, Hei-VAP Advantage, Hei-VAP Precision, Hei-VAP Core, Hei-VAP Expert, Hei-VAP Ultimate, Heizbad Hei-VAP (Heating bin)

Tested according to: UL 61010-1:2012/R:2019-07
 CSA C22.2 No. 61010-1:2012/A1:2018-11
 UL 61010-2-010:2019
 CSA C22.2 No. 61010-2-010:2019

Parameters:

Models: Hei-VAP Value, Hei-VAP Value Digital, Hei-VAP Advantage, Hei-VAP Precision, Hei-VAP Core, Hei-VAP Expert, Hei-VAP Ultimate

Rated voltage: 115 or 230 VAC
 Rated frequency: 50/60 Hz
 Rated power: Max. 1400 W
 Protection class: I

Model: Heizbad Hei-VAP (Heating bin)

Rated voltage: 115 or 230 VAC
 Rated frequency: 50/60 Hz
 Rated power: Max.1300 W
 Protection class: I